



**NMS Labs**

**CONFIDENTIAL**

200 Welsh Road, Horsham, PA 19044-2208  
Phone: (215) 657-4900 Fax: (215) 657-2972  
e-mail: nms@nmslabs.com

Robert A Middleberg, PhD, F-ABFT, DABCC-TC, Laboratory Director

**Demo Report**

**Report Issued** 07/28/2014 08:52  
**Last Report Issued** 06/07/2014 09:25

**88888**  
Clinical Example Report  
Attn: IT Department  
200 Welsh Road  
Horsham, PA 19044-2208

**Patient Name** 1290R-POS  
**Patient ID** 1290R-POS  
**Chain** 14001408  
**Age** Not Given **DOB** Not Given  
**Gender** Not Given  
**Workorder** 14001408  
**Received** 06/04/2014 12:58

**Sample ID** 14001408-001  
**Matrix** RBCs  
**Patient Name** 1290R-POS  
**Patient ID** 1290R-POS  
**Container Type** Clear vial

**Collect Dt/Tm** Not Given  
**Source** Not Given

**Approx Vol/Weight** Not Given

**Receipt Notes** None Entered

Analysis and Comments	Result	Units	Reporting Limit	Notes
-----------------------	--------	-------	-----------------	-------

**1290R Cobalt, RBCs**

Analysis by Inductively Coupled Plasma/Mass Spectrometry (ICP/MS)

Cobalt	11	mcg/L	1.1	ELEVATED
--------	----	-------	-----	----------

NMS Labs derived data:  
2.5th - 97.5th percentile range is  
<1.0 to 1.5 mcg/L (n=912).

The RBC sample used for analysis was measured by weight and multiplied by the density of human RBC (1.10 g/mL) to obtain mcg/L units.