



NMS Labs

CONFIDENTIAL

200 Welsh Road, Horsham, PA 19044-2208
Phone: (215) 657-4900 Fax: (215) 657-2972
e-mail: nms@nmslabs.com

Robert A. Middleberg, PhD, F-ABFT, DABCC-TC, Laboratory Director

Demo Report

Report Issued 03/30/2020 11:14
Last Report Issued 10/02/2007 11:40

88888
Clinical Example Report
Attn: Example Reports
200 Welsh Road
Horsham, PA 19044

Patient Name 0264LI-POS
Patient ID 0264LI-POS
Chain 1
Age Not Given **DOB** Not Given
Gender Not Given
Workorder 07001684
Received 09/28/2007 06:12

Sample ID 07001684-001
Matrix Liquid
Patient Name 0264LI-POS
Patient ID 0264LI-POS
Container Type Clear vial

Collect Dt/Tm Not Given
Source Not Given

Approx Vol/Weight Not Given

Receipt Notes None Entered

Analysis and Comments	Result	Units	Reporting Limit	Notes
-----------------------	--------	-------	-----------------	-------

0264LI Aluminum, Liquid

Analysis by Inductively Coupled Plasma/Mass Spectrometry (ICP/MS)

Aluminum	5.0	mg/L	0.0050	
No reference data available.				

Specimens for elemental testing should be collected in certified metal-free containers. Elevated results for elemental testing may be caused by environmental contamination at the time of specimen collection and should be interpreted accordingly. It is recommended that unexpected elevated results be verified by testing another specimen.

This test was developed and its performance characteristics determined by NMS Labs. It has not been cleared or approved by the US Food and Drug Administration.