



NMS Labs

CONFIDENTIAL

200 Welsh Road, Horsham, PA 19044-2208
Phone: (215) 657-4900 Fax: (215) 657-2972
e-mail: nms@nmslabs.com

Robert A. Middleberg, PhD, F-ABFT, DABCC-TC, Laboratory Director

Demo Report

Report Issued 04/03/2020 11:28
Last Report Issued 08/07/2018 07:52

88888
Clinical Example Report
Attn: Example Reports
200 Welsh Road
Horsham, PA 19044

Patient Name 2406R
Patient ID 2406R
Chain 18000989
Age Not Given DOB Not Given
Gender Not Given
Workorder 18000989
Received 08/06/2018 14:19

Sample ID 18000989-001
Matrix RBCs
Patient Name 2406R
Patient ID 2406R
Container Type Clear vial

Collect Dt/Tm Not Given
Source Not Given

Approx Vol/Weight Not Given

Receipt Notes None Entered

Table with 5 columns: Analysis and Comments, Result, Units, Reporting Limit, Notes

2406R Indium, RBCs

Analysis by Inductively Coupled Plasma/Mass Spectrometry (ICP/MS)

Table with 4 columns: Indium, None Detected, mcg/L, 0.55. Includes notes: No reference data available. The RBC sample used for analysis was measured by weight and multiplied by the density of human RBC (1.10 g/mL) Not for clinical diagnostic purposes.

Specimens for elemental testing should be collected in certified metal-free containers. Elevated results for elemental testing may be caused by environmental contamination at the time of specimen collection and should be interpreted accordingly. It is recommended that unexpected elevated results be verified by testing another specimen.

This test was developed and its performance characteristics determined by NMS Labs. It has not been cleared or approved by the US Food and Drug Administration.