NMS Labs
200 Welsh Road, Horsham, PA 19044-2208
Phone: (215) 657-4900 Fax: (215) 657-2972
e-mail: nms@nmslabs.com
Robert A. Middleberg, PhD, F-ABFT, DABCC-TC, Laboratory Director

CONFIDENTIAL

Demo Report

Report Issued 04/03/2020 11:54
Last Report Issued 04/07/2017 19:32

88888
Clinical Example Report
Attn: Example Reports
200 Welsh Road
Horsham, PA 19044

Patient Name 3140N
Patient ID 3140N
Chain 17000713
Age Not Given **DOB** Not Given
Gender Not Given
Workorder 17000713

Received 04/06/2017 08:48

Sample ID 17000713-001
Matrix Nails
Patient Name 3140N
Patient ID 3140N
Container Type Clear vial

Collect Dt/Tm Not Given
Source Not Given

Approx Vol/Weight Not Given

Receipt Notes None Entered

Analysis and Comments	Result	Units	Reporting Limit	Notes
3140N Nickel, Nails				
Analysis by Inductively Coupled Plasma/Mass Spectrometry (ICP/MS)				
Nickel	None Detected	mcg/g	0.080	

No reference data available.
Not for clinical diagnostic purposes.

Specimens for elemental testing should be collected in certified metal-free containers. Elevated results for elemental testing may be caused by environmental contamination at the time of specimen collection and should be interpreted accordingly. It is recommended that unexpected elevated results be verified by testing another specimen.

This test was developed and its performance characteristics determined by NMS Labs. It has not been cleared or approved by the US Food and Drug Administration.